
MODULE NAME: MNC

REVISION LEVEL: A

MODULE TYPE: IOMOD

I.D. NUM.: 94

DEVICES TESTED: MNCKW (CLOCK) MINC MODULE

DEFAULT PARAMETERS: DVA-171020 VCT-440 BR1-4
BR2-0 DVC-1 SR1-0

REQUIRED PARAMETERS: NONE

DEVICE SETUP: THE FRONT PANEL SWITCHES MUST SELECT
THE SCHMITT TRIGGER INPUT. (PULL OUT
"ST1" AND "ST2" AND ROTATE TO THE END.)

MEANING OF SR1: SELECTS TEST FREQUENCIES. A '1'
DISABLES TESTING.

BIT0 = 1MHZ
BIT1 = 100KHZ
BIT2 = 10KHZ
BIT3 = 1KHZ
BIT4 = 100HZ
BIT5 = RANDOM FREQUENCY (OVERRIDES
DESELECTION OF ANY FREQUENCY)
